

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

SEP 10 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130301 ✓
Name	Kyle Carrington	ID#	5336 Date 9-4-09 ✓
A	Agency	UAFAO	Phone # 474-7721
Instrument Location		612 Yukon Drive Fairbanks, AK 99775 ✓	
Alco S/N		x301583 ✓	Target Value .084 ✓ High Pressure 1100 PSI
B			
Alco Test Values		.087 ✓	.087 ✓
		1 st Alco	2 nd Alco
Signature		Kyle S. Carrington ✓	
(OVER)			

8MB
9/22/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of Oct, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

SEPTEMBER 04, 2009

OPERATOR'S NAME:
CARRINGTON, KYLE
OPERATOR'S NUMBER: 5336
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:

I

O.L. #: 0

DEPT/AGENCY: FBK7

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .004

ALCO S/N: X301583

— BREATH ANALYSIS —

.004 ADJUSTED FOR 29.70 in
ALCO TARGET .003 08:52
BLANK TEST .000 08:53
INTERNAL STANDARD VERIFIED 08:53
ALCO TU 29.70 in .007 08:54
BLANK TEST .000 08:55
SUBJECT SAMPLE .000 08:55
BLANK TEST .000 08:56
ALCO TU 29.70 in .007 08:56
BLANK TEST .000 08:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

SEPTEMBER 04, 2009

TIME 08:50

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
METERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 40c
BAROMETER: 29.70 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJ KLMNOPQRSTUVWXYZ\]^_`abcdefg h i j k l m n o
p q r s t u v w x y z { | } ~